



Certificate of Compliance

Certificate: 1730355 (LR 80529)

Master Contract: 157918

Project: 2710384

Date Issued: July 21, 2014

Issued to: BW Technologies by Honeywell

2840 2nd Ave SE
Calgary, AB T2A 7X9
Canada
Attention: Steve Uliasz

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Zahra Amíni

Issued by: Zahra Amíni

PRODUCTS

CLASS 4828 01 - SIGNAL APPLIANCES - - Combustible Gas Detection Instruments-For Hazardous Locations

CLASS 4828 81 - SIGNAL APPLIANCES-Combustible Gas Detection Instruments - For Hazardous Location-Certified to U.S. Standards

CLASS 4828 82 - SIGNAL APPLIANCES-Toxic Gas Detection Instruments - For Hazardous Locations. Certified to U.S. Standards

CLASS 4828 02 - SIGNAL APPLIANCES - Toxic Gas Detection Instruments - For Hazardous Locations

Class I, Division 1, Groups A, B, C and D

GasAlertMicroClip XT Portable Gas Detector, MC2-XWHM-a-bb-cc Temperature Code T4 @ $-20^{\circ}\text{C} \leq T_a \leq +58^{\circ}\text{C}$; Multi-gas detectors for monitoring up to four gasses simultaneously and continuously i.e. Oxygen [deficiency/enrichment], Combustibles, Carbon Monoxide and Hydrogen sulphide ; Intrinsically Safe when powered by Narada Lithium Polymer battery Part Number: NL503759.

GasAlertMicroClip XL Portable Gas Detector, MCXL-XWHM-a-bb Temperature Code T4 @ $-20^{\circ}\text{C} \leq T_a \leq +50^{\circ}\text{C}$; Multi-gas detectors for monitoring up to four gasses simultaneously and continuously i.e. Oxygen



Certificate: 1730355 (LR 80529)

Master Contract: 157918

Project: 2710384

Date Issued: July 21, 2014

[deficiency/enrichment], Combustibles, Carbon Monoxide and Hydrogen sulphide ; Intrinsically Safe when powered by Narada Lithium Polymer battery Part Number: NLP883759LT20.

APPLICABLE REQUIREMENTS

CSA Standard C22.2 No. 0-2010 - General Requirements Canadian Electrical Code Part II.

CSA Standard C22.2 No. 152-M1984- Combustible Gas Detection Instruments

CSA Standard C22.2 No. 157-92 - Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations

UL Standard 913, Seventh Edition - Intrinsically Safe Apparatus and Associated Apparatus for use in Class I, II, III, Division 1, Hazardous (Classified) Locations

ANSI/ISA-12.13.01-2000 - Performance Requirements for Combustible Gas Detectors